KLY® LUBRICATING JELLY

EN

PRODUCT NAME: KLY® LUBRICATING JELLY
COMPOSITION: Deionized water, Natural Glycerin, MonopropylenGlycol, Hydroxy Ethyl Cellulose, Methyl Hydroxybenzoate, Citric Acid
INTENDED USE: This gel is used in endoscopy, colonoscopy, gynecological examinations, rectal endoscopic, catheter operations, also used for vaginal lubrication.
INSTRUCTION FOR USE: To Open The Tube: Remove and reverse the cap then puncture the tube with it. Apply Lubricating Jelly to the gloves or the instrument in sufficient amount so that it forms a film over the area that comes in contact with patient. Wash off after use.
To Open The Sachet: Open carefully from marked area.
SPECIAL PRECAUTIONS: Precautions: If container is damaged or opened, the Product is not guaranteed.
CONTRAINDICATION: There is no known contraindication
OVERDOSAGE AND SIDE EFFECTS: There is no known symptoms for overdose usage.
SENSITIVITY TO GLYCERIN: Glycerin has emollient properties. Precautions Shall be taken when using the gel in people with sensitivity to glycerin.
IDENTIFICATION: Transparent, water-soluble, greaseless and jelly.
PRESENTATION: Tubes Containing either: 42 g, 50 g, 82 g, 113,4 g, 100 g, 250 g, 1000 g
SACHETS CONTAINING EITHER: 2,5 g, 3 g, 3,5 g, 4 g, 4,5 g, 5 g, 10 g
STORAGE INSTRUCTIONS: Store between 5-30°C
KEEP OUT OF REACH OF CHILDREN

TRADEMARK / KLY

Paket hasarlı ise kullanmayınız *Do not use if package is damaged* يجب عدم استعمال المنتج اذا كانت العبوة متضررة	Güneş ışığından uzak tutunuz *Keep away from sunlight* يجب المحافظة على المنتج بعيدا عن أشعة الشمس	Kuru ortamda saklayınız *Keep Dry* يجب المحافظة على المنتج في وسط جاف
Saklama Sıcaklığı *Storage Temperature* درجة حرارة التخزين	Dikkat *Caution* انتباه	Katalog Numarası *Catalogue Number* رقم الكتالوج
Kullanım kılavuzuna bakınız *Consult instructions for use* يجب مراعاة تعليمات الإستخدام	Üretici *Manufacturer* المصنع	"LOT" numarası *Batch Code* رقم الدفعة
Üretim Tarihi *Date of Manufacture* تاريخ الانتاج	Son kullanma tarihi (gg/ yıl) *Expiration Date (Months/ Year)* تاريخ انتهاء الصلاحية (شهور/ السنة)	Tek kullanımlıktır. *Do not re-use.* وحد الاستعمال
Nonsteril / Non-sterile غير معقم	Tıbbi cihaz *Medical Device* جهاز طبي	

Rev:10/23.02.2023

Turkuaz Sağlık Hizmetleri, Medikal, Temizlik Kimyasal Ürünler San. ve Tic. A.Ş.
Akçaburgaz Mah. Muhsin Yazıcıoğlu Cad. No:45/5 34522 Esenyurt / İstanbul / Türkiye
T +90 212 428 6848 Pbx **F** +90 212 428 6853
info@turkuazsaglik.com.tr / www.**turkuazsaglik**.com.tr

CE 2292

Made in Turkey
Discover the potential

PROSPEKTÜS/IFU

Tarih: 23 / 02 / 2023

ÜRÜN KODU	: 150.06.05.0582		
ÜRÜN İSMİ	: KLY NON-STERIL ORTAK PROSPEKTUS - YENI ADRES - REV.10		
MALZEME CİNSİ	: I.HAMUR 70 GR		
GİRECEĞİ KUTU ÖLÇÜSÜ	: 135x40x30mm	KATLAMA	: OLACAK
RENK	: BLACK	ÖLÇÜ	: 285X120 MM

AÇIKLAMA / NOT

Turkuaz Sağlık Hizmetleri, Medikal, Temizlik Kimyasal Ürünler San. ve Tic. A.Ş.
Akçaburgaz Mah. Muhsin Yazıcıoğlu Cad. No:45/5 34522 Esenyurt / İstanbul / Türkiye
T +90 212 428 6848 Pbx **F** +90 212 428 6853 **info@**turkuazsaglik.com.tr / www.**turkuazsaglik**.com.tr