CONNECTION

Cavo Collegamento Elettrodo Neutro
Connection Cable for Neutral Electrode

LED SpA
Via Selciatella 40
04011 Aprilia (LT) Italy

REF 00404.08_S

LOT 000000

UDI (01) 08033638324547
(17) YYMMDD
(10) 000000

IT YYYY-MM-DD

YYYY-MM-DD

x 1

MD NON STERILE CE 0051 0°C + 50°C 20% 80% 04 PE-LD

INTENDED USE:

Devices destined to be connected, for temporary use, to high frequency electrosurgical units for surgical operations in which cutting and / or coagulation of soft tissues is required, with a monopolar and / or bipolar technique, for minor surgery and / or major in open and / or intra-operative percutaneous and / or endoscopic and / or laparoscopic. See attached table.

WARNING:

The following precautions reduce the risk of accidental burnings.

- The product must be used with ESU and/or accessories comply with current regulations.
- To use the ESU and other necessary accessories refer to instructions of the individual elements.
- Do not let the wires come into direct contact with the patient's skin, a capacitive effect could cause burns.
- When removing it, do not pull the cable directly but act on the connector.
- The user and/or the patient shall be informed of the need to report any serious incident that has occurred in relation to the device, the manufacturer and the competent authority of the Member State where the user is established.

CLEANING:

Wash with running water immediately after the operation, then plunge for around 45 minutes in a solution 2% of a normal disinfection product.

COMPATIBLE WITH DISPOSABLE NEUTRAL ELECTRODE /FLEX /FLEX S /5365 /6429





ed. 05-2025

IA00404.08c_EN-IT