CONNECTION

Cavo Collegamento Elettrodo Neutro
Connection Cable for Neutral Electrode

LED SpA
Via Selciatella 40
04011 Aprilia (LT) Italy

REF 00404.08_S

LOT 000000

UDI (01) 08033638324547
(17) YYMMDD
(10) 000000

IT YYYY-MM-DD

YYYY-MM-DD

x 1

MD NON STERILE CE 0051 0°C + 50°C 20% 80% 04 PE-LD

INTENDED USE:

Devices destined to be connected, for temporary use, to high frequency electrosurgical units for surgical operations in which cutting and / or coagulation of soft tissues is required, with a monopolar and / or bipolar technique, for minor surgery and / or major in open and / or intra-operative percutaneous and / or endoscopic and / or laparoscopic. See attached table.

WARNING:


The following precautions reduce the risk of accidental burnings.


- The product must be used with ESU and/or accessories comply with current regulations.
- To use the ESU and other necessary accessories refer to instructions of the individual elements.
- Do not let the wires come into direct contact with the patient's skin, a capacitive effect could cause burns.
- When removing it, do not pull the cable directly but act on the connector.
- The user and/or the patient shall be informed of the need to report any serious incident that has occurred in relation to the device, the manufacturer and the competent authority of the Member State where the user is established.

CLEANING:

Wash with running water immediately after the operation, then plunge for around 45 minutes in a solution 2% of a normal disinfection product.

COMPATIBLE WITH DISPOSABLE NEUTRAL ELECTRODE /FLEX /FLEX S /5365 /6429





ed. 05-2025

IA00404.08c_EN-IT

DESTINAZIONE D'USO:

Dispositivi destinati ad essere collegati, per uso temporaneo, ad apparecchi di elettrochirurgia ad alta frequenza per interventi chirurgici nei quali è richiesto il taglio e/o la coagulazione di tessuti molli con tecnica monopolare e/o bipolare, per chirurgia minore e/o maggiore in aperto e / o intraoperatorio percutaneo e / o endoscopico e / o laparoscopico.

AVVERTENZE:

Le seguenti precauzioni hanno lo scopo di ridurre il rischio di ustioni accidentali.

- Il prodottto deve essere utilizzato con ESU e/o accessori conformi alle normative vigenti.
- Per l'uso delle ESU e degli altri accessori necessari riferirsi alle istruzioni dei singoli elementi.
- Evitare che i cavi entrino in contatto diretto con la pelle del paziente, un effetto capacitivo potrebbe provocare bruciature e/o ustioni.
- Nel rimuovere non tirare direttamente il cavo ma agire sul previsto connettore.
- In caso di necessità all'utilizzatore e/o paziente devono segnalare qualsiasi incidente grave verificatosi in relazione al dispositivo al fabbricante e all'autorità competente dello stato membro in cui l'utilizzatore e/o il paziente è stabilito.

PULIZIA:

Lavare con acqua corrente subito dopo l'intervento e poi immergere per 45 minuti in una soluzione al 2% di un nirmale prodotto disinfettante.

COMPATIBILE CON ELETTRODI NEUTRI MONOUSO /FLEX /FLEX S /5365 /6429