GIMA

PROFESSIONAL MEDICAL PRODUCTS

SEDIA PORTANTINA PER SCALE VALUE - 4 RUOTE
VALUE STAIR CHAIR - 4 WHEELS

ITALIANO

È necessario segnalare qualsiasi incidente grave verificatosi in relazione al dispositivo medico da noi fornito al fabbricante e all'autorità competente dello Stato membro in cui si ha sede.

I. Parametri tecnici:

1. Dimensioni da piegata: 92 x 52 x 16 cm
2. Piegata in sedia a rotelle: *altezza dello schienale:* 91cm - larghezza: 50 cm
   seduta: altezza: 49 cm - larghezza: 50 cm
3. Peso netto: 8 kgs

II. Metodo operativo:

1. Questo modello è dotato di una lega di alluminio di alta qualità e di pelle Oxford. Per rendere più sicuro il trasporto del paziente, in dotazione ci sono delle cinture.
2. Aprire la barella per sedia a rotelle e tirare la parte 4 e spingerla verso il basso, dopo averla regolata, premere verso il basso la parte 2, fino al punto stabilito, la sedia a rotelle si bloccherà automaticamente.
3. Premere il pulsante sulla parte 2, la maniglia può essere tirata fuori e spinta indietro a piacimento. Premere la parte 3, è possibile chiudere la maniglia nella parte posteriore.
4. Per piegare la barella per sedia a rotelle, sollevare la parte 1, salire su una delle guide poste davanti alle rotelle, nel frattempo sollevare la parte 4 e quindi piegare la barella.

III. Modalità di montaggio:

Questo modello si usa principalmente in aree caratterizzate da un traffico o accesso difficile, come corridoi stretti, cancelli, cortili o alcune zone di passaggio.

IV. Manutenzione:

1. Mantenere la pulizia, antisepsi inclusa.
2. Verificare la presenza di eventuali parti della barella allentate.

V. Attenzione:

Durante il trasporto del paziente, assicurarsi che ogni parte funzioni bene e allacciarsi la cintura.

VI. Conservazione e trasporto:

1. Questo prodotto deve essere conservato in un ambiente asciutto e non soggetto ad erosione.
2. Questo prodotto può essere trasportato da un veicolo di trasporto ordinario.

VII. Imballaggio e accessori:

1. Questo prodotto è imballato completamente in una confezione singola.
2. Accessori: certificato di idoneità e foglio catarifrangente.

	Fabbricante		Conservare in luogo fresco ed asciutto	CE	Dispositivo medico conforme al regolamento (UE) 2017/745
REF	Codice prodotto		Conservare al riparo dalla luce solare		Attenzione: Leggere e seguire attentamente le istruzioni (avvertenze) per l'uso
LOT	Numero di lotto	MD	Dispositivo medico	EC REP	Rappresentante autorizzato nella Comunità europea
	Leggere le istruzioni per l'uso		Data di fabbricazione		Importato da:
UDI	Identificativo Unico Dispositivo				

CONDIZIONI DI GARANZIA GIMA

Si applica la garanzia B2B standard Gima di 12 mesi.

ENGLISH

All serious accidents concerning the medical device supplied by us must be reported to the manufacturer and competent authority of the member state where your registered office is located.

I. Technical Parameter:

1. fold Size: 92 x 52 x 16 cm
2. fold into wheelchair: *backrest height*: 91 cm - width: 50 cm
   seat: height: 49 cm - width: 50 cm
3. Net weight: 8 kgs

II. Operating Method:

1. This model is equipped with high quality aluminum alloy as well as Oxford leather. To make the carrying of patient more safety, belts are equipped.
2. Open the wheelchair stretcher and pull part 4 and press it down,
3. after settled, then press down part 2, to the standard point, the wheelchair will become locked automatically.
4. Press the button on part 2, the length of the handle can be pulling out and pushing back at will. Press part 3, the handle at the back part can be folded.
5. When folding the wheelchair stretcher, lift part 1, step on one of the rails in front of the castors, meanwhile, lift part 4, and the stretcher can be folded.

III. Fitting Scope:

This model is mainly used in areas with difficult traffic or entry such as narrow halls, gateways, cloisters or several turnings.

IV. Maintenance:

1. Keep it clean, antisepsis included.
2. Check whether any part of the stretcher become loosens.

V. Attention:

When carrying the patient, insure every part works well and fasten the belt.

VI. Storage & transportation:

1. This product must be stored in the environment of dry and no erode.
2. This product can be transported by every day transporting vehicle.

III. Packing and accessories:

1. This product is packed in a carton as a whole.
2. Accessories: eligibility certificate as well as illumination paper.

	Manufacturer		Keep in a cool, dry place	CE	Medical Device compliant with Regulation (EU) 2017/745
REF	Product code		Keep away from sunlight		Caution: read instructions (warnings) carefully
LOT	Lot number	MD	Medical Device	EC REP	Authorized representative in the European community
	Consult instructions for use		Date of manufacture		Imported by:
UDI	Unique Device identifier				

GIMA WARRANTY TERMS

The Gima 12-month standard B2B warranty applies.

Zhangjiagang Tengda Machinery Manufacturer Co.,Ltd
No.999 Hongqi Road, Changyinsha
Modern Agricultural Demonstration Area,
Zhangjiagang City, Jiangsu, 215623, China
Made in China

Gima 34073

REF **TD010114A**

EC REP SUNGO Europe B.V.
Fascinatio Boulevard 522, Unit 1.7, 2909VA Capelle aan den Ijssel, The Netherlands

Gima S.p.A.
Via Marconi, 1 - 20060 Gessate (MI) Italy
gima@gimaitaly.com - export@gimaitaly.com
www.gimaitaly.com

M334073-M-Rev3-0526