GIMA

PROFESSIONAL MEDICAL PRODUCTS

SUPPORTO A ROTELLE
ROLLING STAND
CHARIOT À ROULETTES
SOPORTE RODANTE
SUPORTE ROLANTE



	REF
GIMA 35127	8030421050
GIMA 35148	8030421049
GIMA 54422	8030421060

Fabbricante / Manufacturer / Fabricant / Fabricante / Fabricante:
Shenzhen Stone Wisdom Technology Co., Ltd.Room 401,
Building 6, Fanshen Industrial Zone, Yangguang Second Road,
Yangguang Community, Xili Street, Nanshan District, Shenzhen
mei@stom-pd.com
MADE IN CHINA

Importato da / Imported by / Importé par / Importado por / Importado por:
Gima S.p.A.
Via Marconi, 1 - 20060 Gessate (MI) Italy
gima@gimaitaly.com - export@gimaitaly.com
www.gimaitaly.com

M35127-M-Rev.102.25

Version: Ver1.2

Date: December 23, 2024

Instructions to user

In order to operate properly and get the best performance, please read these instructions carefully before using this rolling stand. This manual describes installation, operation, maintenance, transportation and storage.

Please keep this manual for future reference!



1. Features



Silent castors with brakes



Basket with handle and hook



Height adjustable



Elevation angle adjustable



Mounting plate lock

2. Intended Use

This product is intended for transferring equipment such as Patient Monitor, ECG Monitor, etc. to help doctors work conveniently and effectively.

3. Cleaning and Disinfection

1. Remove the monitor before cleaning the rolling stand.

2. It is recommended to clean the rolling regularly. Only a non-corrosive cleanser is permitted. Dilute cleansing products as per the manufacturers instructions.

3. Use a cloth with alcohol or your choice of cleanser to wipe the surface of the rolling stand, and dry it with a clean cloth or leave to air-dry. all'aria.

4. Specification

Dimension: **503*461*H924-1074 mm (height adjustable)**

Weight (including package): **12 kg**

Material:

Top pole: **Stainless Steel**

Lower pole : **Aluminium Alloy**

Base: **Aluminium Alloy**

Basket: **Iron**

Overall loading-bearing: **30 kg**

5. Transportation and storage

1. Transportation

The rolling stand should be left in original packaging when being transported and kept dry.

2. Storage conditions:

The rolling stand should be kept in a dry and well ventilated room free from dust and corrosive gases.

If the stand is being stored in original packaging, stacking height should not exceed four layers, warehouse should be room temperature.

6. Packing List

Base kit *1

Mounting plate*1

Support pillar *1

Castor *4

Installation tool *1

7. Troubleshooting

Problem	Analysis	Solution
Support pillar shakes	Fixing screws on the base may be loose	Check the screws and tighten if required

Elevation angle can be adjusted	Adjusting screws on the handle are too loose / tight	Adjust the tightness according to the requirements


8. Installation

Part List





No	Part Name	Qty
1	Mounting plate	1
2	Storage Basket	1
3	Support Pillar	1
4	Base	1
5	Castor	4
6	Screw M4-10/12	8
7	Screw M5-16	5
8	Screw M6-25	4
9	3 mm spanner	1
10	4 mm spanner	1

Step 1:
Fix four castors into the related holes on the base.




Step 2:
Insert the support pillar into the base and secure with M6-25 screws.






Step 3:
Assemble the storage basket, and secure with M4-10 and M4-12 screws according to the following pictures



Step 4:
Fix the bracket into the support pillar and secure with M5-16 screws



Legend of symbols

	Fragile, handle with care	REF	Product code
LOT	Lot number		Keep away from sunlight
	Keep in a cool, dry place		

GIMA WARRANTY TERMS
The Gima 12-month standard B2B warranty applies..